

The benefit of using an **Independent** Quality Control:

Case Studies

A real life/real Lab experience 2014 to 2019

Labroots

Bio-Rad Webinar

Oct-21-2020

Abubaker Yagoot ابوبكر ياقوت
Supervisor Biochemistry Section
M.N.G.H.A. King Abdulaziz Medical City - Jeddah

Disclaimer

- * No Financial conflicts to disclose.
- * Information in this presentation and examples are for Educational purposes only.
- * No product advice or endorsement.
- * Not promoting any service or product.

Learning Objectives

Participants will be able to:

1. Identify at least 2 benefits Independent Q.C. can provide when troubleshooting an out-of-control event.
2. Recognize at least 2 advantages an interlaboratory comparison program can provide.

Agenda

4 case study examples

- 1. Unity** Monthly Peer comparison **revealed marginal performance.**
Example “**GGT**” calibration factor.
- 2. Unity** Monthly Peer comparison **provided confidence.**
Example “**Vancomycin**”
- 3. Unity** Monthly Peer comparison (19-02-2019), 2 level controls outside the acceptable 2 SDI of Peer group (69 Labs, 2434 points).
Example “**Micro albumin Urine Albumin**” Calibrator.
- 4. Free T3.** New 6 point calibrator and new Assay file. **Stop test.**

Interpretation

OCNet Monthly Report

Unity™ Laboratory Performance Overview
Immunoassay Plus • Lot 40930 • Exp 28–Feb–2019

November 2017 • Lab 681268

CLINICAL CHEMISTRY


SDI=How many S.D away from the over all mean

Bias

Precision

Suggested Interpretation

The following interpretations should be used as guidelines only. Each laboratory should determine whether a review of bias and/or precision is necessary and whether corrective actions may be required.



A Acceptable Performance.

B Acceptable to Marginal Performance – May indicate the need to investigate test system bias or precision.

C Marginal Performance – Corrective action may be necessary.

- Your laboratory coordinates for bias and precision based on your peer group.
- Your laboratory coordinates for bias and precision based on the method group.

Notes SDI is a statistical estimate of bias.

CVR is a ratio of monthly laboratory CV to monthly peer group CV and is an estimate of precision.

Formulas




Laboratory Performance Overview
Immunoassay Plus • Lot 40930 • Exp 28–Feb–2019

November 2017 • Lab 681268
CLINICAL CHEMISTRY

Suggested Interpretation

The following interpretations should be used as guidelines only. Each laboratory should determine whether a review of bias and/or precision is necessary and whether corrective actions may be required.



- A Acceptable Performance.
 - B Acceptable to Marginal Performance – May indicate the need to investigate test system bias or precision.
 - C Marginal Performance – Corrective action may be necessary.
- Your laboratory coordinates for bias and precision based on your peer group.
 - Your laboratory coordinates for bias and precision based on the method group.
- Notes** SDI is a statistical estimate of bias.
CVR is a ratio of monthly laboratory CV to monthly peer group CV and is an estimate of precision.

S.D.I and C.V.R


$$\text{S.D.I} = \text{Standard Deviation Index} = \frac{\text{Lab Mean} - \text{Peer Group Mean}}{\text{Peer group S.D}} = \text{Bias (how many S.D away from the overall Mean)}$$

$$\text{C.V.R} = \text{C.V. Ratio} = \frac{\text{Lab C.V (monthly)}}{\text{Peer C.V (monthly)}} = \text{Precision}$$

Case Study # 1


G.G.T (Gamma-Glutamyl Transferase)

Unity[™] Laboratory Performance Overview Assayed Chemistry • Lot 26410 • Exp 31-Jan-2019




Unity "Laboratory Comparison Report"

GGT (October 2017)



Unity “Monthly Bias & Imprecision Histogram”

GGT (December 2016)



Unity "Bias & Imprecision Histogram"

GGT (October 2017)





Unity

Bias & Imprecision Histogram

Assayed Chemistry • Lot 26410 • Exp 31-Jan-2019

October 2017 • Lab 681271




CLINICAL CHEMISTRY



Unity “Bias & Precision Histogram”

GGT (September 2016)

Unity Bias & Imprecision Histogram
 Assayed Chemistry • Lot 14490 • Exp 30–Nov–2017
 September 2016 • Lab 681271
 CLINICAL CHEMISTRY



Product Information

December-2017

“Adjust the GGT calibration Factor using a standardized alignment solution.”



Product Information

Date Issued September 19, 2014

Received on (20-November-2017)

Product

Product	List Number (LN)
Gamma-Glutamyl Transferase (GGT) Reagent	7D65


Reason

We are pleased to announce a new optional GGT alignment procedure for improved ARCHITECT cSystems instrument-to-instrument GGT assay results. This procedure adjusts the GGT calibration factor using a standardized alignment solution and will be available 4th quarter 2014. The procedure provides:

- Better alignment of multiple instruments within a single laboratory
- Better agreement between laboratories across a network


The **disadvantage** of not having a means of comparing your Q.C results with your Peer group, example Unity.

Levy Jennings chart from Instrument looks perfect. **Before changing the factor.**



The **disadvantage** of not having a means of comparing your Q.C results with your Peer group, example Unity.

Levey Jennings chart from Instrument looks perfect. [After changing the factor.](#)




The Advantage of having a means of comparing your Q.C with your Peer group, example Unity.

Factor changed on this day, (26-12-2017).

Readings of the same control Before and after the Factor change showing an 18% falsely elevated result.

Which explains the years of continuous +ve bias for Q.C as well as patient result.




Problem solved for the Lab

Any further action required??

Factor changed on this day
Time completed: 26.12.2017 / 13:17
Value: 158 U/L

No patient harm was done.

GGT showed a Marginal performance, but with in 2 S.D.



communicated to Westgard QC whenever this practice is implemented. However, it is not acceptable for the laboratory to apply less demanding quality requirements or implement less stringent QC procedures than those suggested by the Westgard Sigma VP. Violations of this type will result in immediate disqualification of the entire laboratory from the program.

For non-verified assays, the QC practices will not be scrutinized by the Westgard Sigma VP.

Conclusion

For the analytes listed in this report, under the circumstances detailed in the report, Westgard QC, Inc. is proud to RE-verify that NGHA Jeddah is achieving the appropriate goals of analytical quality performance for its chemistry assays.

A brief description of this updated verification will be available online for the duration of the verification period at <http://www.westgard.com/sigmavp-ngha-jeddah.htm>

Sten Westgard

Sten Westgard
Director, Client Services and IT,
Westgard QC, Inc.
7614 Gray Fox Trail
Madison WI 53717

Unity


September 2017 • Lab 681271

CLINICAL CHEMISTRY
P. O. BOX 9515
Jeddah Jeddah
Attention: KING ABDULAZIZ MEDICAL CITY, WIR
PATHOLOGY

Monthly Evaluation
Assayed Chemistry • Lot 26410 • Exp 31-Jan-2019

Please review your QC reports for September 2017.

Your tests are all within established parameters.



The magnitude of Risk

If there was no Peer comparison

The direct effect of all that to the patient

Work Load Statistics of GGT from 2014 to 2017

Year	source of data	# of Samples	Number of Patients at 3.3 samples per patient (source) AinIQ-BIS
2014	Estimate	70,000	
2015	Estimate	70,000	
2016	CERNER	115,601	
2017	CERNER	73,741	
Total		329,342	99,800 = 100,000 patients since 2014 in Jeddah only

1

2

3

? # of Patients

Cost of Good Quality

Vs

Cost of Poor Quality

1 vial per 3 days, 1 kit = 12 vials = SAR 1000

1 kit of Q.C-1 & Q.C-2, 2 kits per month = SAR 2000

In 3 Years 2000 x 36 = **SAR 72,000 (\$ 19,200)**

SAR 100,000,000
(\$26,666,666)

Cost of Quality



- CLSI Report QMS20-R. 2014 (understanding the cost of Quality in the Laboratory)
- Cost of Quality for Return on Investment (ROI)
- Cost of Good Quality (prevention/appraisal/upfront cost)
- Cost of Poor Quality (internal failure/external failure cost)
- Hard cost vs soft cost
- Cost saving Vs cost avoidance , templates/Forms/Charts

$$\text{R.O.I} = \frac{\text{Amount Gained} - \text{Amount spent}}{\text{Amount Spent}} \times 100$$

Laboratory Performance Overview (GGT)

Before changing the Factor



After changing the Factor



Laboratory Comparison Report (GGT)


Before changing the Factor

After changing the Factor




Bias imprecision (GGT)

Before changing the Factor



After changing the Factor



Learning from mistakes (GGT)

What are the chances of this will:

1. Happen to your Lab?

or

2. Happen to our Lab again?

Why IFCC factor?

† Due to differences in instrument systems and unit configuration, the GGT insert package parameter is available in ARCHITECT Software version 7.00 and later.

†† The c8000 Primary Wavelength is 412 nm; the c16000 and c4000 Primary Wavelength is 416 nm.

††† The calibration factor for c8000 is 8077 (IFCC factor = 7418); the calibration factor for c16000 is 8750 (IFCC factor = 8036). For c4000, the calibration factor is 9116 (IFCC factor = 8372).

††† The linear low value (Low-Linearity) is LOQ rounded up to the number of decimal places defined in the decimal places parameter field.

GGT insert package

Enzyme Reagent Master List

Deleted codes		New/Updated codes	
None			
C.A.P. (C-B 2020) Kit Instruction			
All reaction temperatures are assumed to be 37 °C. If your instrument uses a temperature other than 37°C, please use code "0010 Other Reagent/ Reaction Temperature, specify on result form."	1915	DiaSys without Pyridoxi-5-phosphate	1091
	1089	Erba	Roche Pseudocholinesterase (CHE) - Acetylthiocholine method
	1342	Formosa Biomedical Technology MeDiPro	1092
	1718	Gilford	Roche Pseudocholinesterase Gen.2 (CHE2) - Butyrylthiocholine method
1045	Abaxis	1641	Global Chemical Diagnostics
1518	Abbott	1925	Hemagen
2052	Abbott IFCC	1173	HI Chem Diagnostics
1526	Abbott Activated ALT/AST with P5P (IFCC)	1100	Horiba ABX
		1046	Roche with Pyridoxal L-5-Phosphate
		1037	Sekisui
		2084	Sekisui LD Lactate to Pyruvate
		1047	Sekisui LD Pyruvate to Lactate

Attachment A

The optional GGT alignment procedure will only be performed by Abbott Service personnel. The alignment procedure should be performed in the laboratory. The current calibration factors in the reagent package insert are:

- c8000 — 8077 (IFCC factor = 7418)
- c16000 — 8750 (IFCC factor = 8036)
- c4000 — 9116 (IFCC factor = 8372)

Product Information from Sept-19-2014

Configure assay parameters – Calibration

- The laboratory must evaluate if QC will require **Product Information**
- In the future, if an update to the GGT parameters is installed via an assay disk or AbbottLink, the laboratory must ensure the instrument-specific calibration factor is re-configured in the GGT parameters.
- No further evaluation of the new calibration factor(s) is needed unless the optics is replaced or a new ARCHITECT cSystem (for running GGT) is installed in your laboratory.

General **Calibration** SmartWash Results Interpretation

Assay: GGT Assay number: 1027 Date: 29.09.2020

Calibration method: Factor Time: 12:51

Factor: 8077.0000 Operator: ADMIN

non IFCC factor is the download default.

General **Calibration**

Assay: GGT Assay number: 1027 Date: 30.09.2020

Calibration method: Factor Time: 15:03

Factor: 7418.0000 *IFCC factor* Operator: ADMIN

Calibrators Volumes Intervals Validity checks

Summary

1- Case # 1 (GGT)

- **Unity** Monthly Report showed marginal Performance. (Continuous Positive bias)
- Investigation - vendor Product Information Letter recommended Adjust Calibration Factor.
- Calibrator Adjusted – Problem Solved.
- **Second wave** of GGT + ve bias, (29-Sept-2020)

Case study # 2

Vancomycin

Feb-2018, Pharmacy called the Lab.



You are giving us....

“higher results than expected of Vancomycin trough levels.”

Checked Q.C results on Instruments

Checked Q.C results and Q.C History on Unity (www.QCNet.com)

Checked recent C.A.P. surveys



Our Controls are good compared to our Peers.




Case study # 2

Vancomycin

Pre-analytical

Feb-2018, Pharmacy, higher results than expected since November-2017


Two Instruments performance with Quality Control Material. (Peers = 75 Labs, Method = 122 Labs)



Feb-2018, Pharmacy, higher results than expected since November-2017

➤ Two Instruments performance with patient samples in the past Two years.

Total numbers of Vancomycin Critical results (Greater than 20.99 ug/ml), phoned to the physician (Jan 2017 to Sept 2018)



Conclusion

- No further Investigation in Analytical phase.
- Investigate Pre-Pre-Analytical. (Pharmacy)


➤ “Up to 75% of Lab errors occur in pre-analytical phase.” ??

What happen ?
Why this Example ?



Feb-2018, Pharmacy, higher results than expected.

Pre-Pre-Analytical Investigation



Summary

2- Case # 2 (Vancomycine)

- Pharmacy queried “Vancomycin trough results lately higher than expected.”
- **Unity** Monthly Report showed Good Q.C Performance compared with Peer group. (Peer/Method = 75/122 Labs) (confidence)
- No further Investigation in Analytical Phase.
- Origin of Problem was found to be Pre-Pre-Analytical. (Medication)

19-02-2019, Unity Real Time (Monthly Evaluation Report)

Urine Chemistry Control lot # 66800 Exp. Date 31-03-2019

- **Two levels of control outside the acceptable Peer 2 SDI. compared to 69 Laboratories Globally (2434 points).**
- **Two Instruments**

Case study # 3

Micro albumin/Urine Albumin



Calibrator

Micro albumin/Urine Albumin

19-02-2019, Unity Real Time (Monthly Evaluation Report)

Finding/Problem

- Urine Chemistry Control lot # 66800 Exp. Date 31-03-2019
- Two levels of control out side the acceptable Peer 2 SDI, compared to 69 Laboratories Globally (2434 points), on Two Instruments.




Micro albumin/Urine Albumin

19-02-2019, Unity Real Time (Monthly Evaluation Report)

Urine Chemistry Control lot # 66800 Exp. Date 31-03-2019

Investigation

- Current in use Micro albumin calibrator (Ref. # 2K98-02)
- Insert package for Micro albumin reagent
- Under Calibration paragraph:- “The Microalbumin assay must be calibrated using the microalbumin calibrator ref # 2K98-03”



Micro albumin/Urine Albumin

19-02-2019, Unity Real Time (Monthly Evaluation Report)

Urine Chemistry Control lot # 66800 Exp. Date 31-03-2019

Solution

- Recalibrated both Instruments with the correct calibrator and run the same control. Result the same as the peer value. See Levy-Jennings graph print out.


Case study # 3

Micro albumin

calibrator


Micro albumin/Urine Albumin

Levy Jennings graph – Instrument print out



Micro albumin/Urine Albumin

- Bio-Rad Notification. Unity Report Available in QCNet.



- E-mail from Bio-Rad on Feb-19-2019 @ 9:04 PM, opened on 20-02-2019
- Investigation done and problem solved for Two instruments 19-02-2019 @ 3:40 PM
- How was the Lab able to do that???
- (Unity Monthly Report printed on 17-Feb-2019 @18:07:51)

Why this Example ?

Summary

3- Case # 3 (Microalbumin / Urine Albumin)

- **Unity** Report for January 2019:
Both the two Controls for the two Instruments outside the Peer range.
- Investigation – As per Reagent Insert Package – needed different version of calibrator than the one in use.
- Calibrated – Problem solved.

Product Information on (September-10-2019)

New assay file, new reagent, new calibrator (6 point)


Immediately after starting the new calibrator:

- Bio-Rad Q.C results were low outside the manufacturer range for Two different lot #s.
- The Two lots of Independent control picked up the problem
- Single / Company Q.C all level on a low side but with in the range.
- The manufacturer control did not pick up the problem
- Rerun Previous C.A.P survey (C-B 2019)
- One C.A.P. sample was out and the rest of the P.T samples all on a negative bias .

Product Information on (September-10-2019) New assay file, new reagent new calibrator (6 point)

➤ Unity Real Time (L.J chart) showing downwards shift

Oct-04-2019, Immunoassay plus **Lot # 40950** First Q.C post 6 point calibration.
(First 6 point calibrator)




➤ Problem solved with new 6 point calibrator.

Product Information on (September-10-2019)

New assay file, new reagent new calibrator (6 point)

Action

- Immediately **stopped** the Assay (Free T3) until the correct Calibrator arrives from outside.




P	ASSAY	CAL STATUS	REMAINING TESTS	REAGENT STATUS
10	AFP_3	Active	81	OK
21	B-hCG STAT	Active	51	OK
9	CEA	Active	55	OK
6	CK-MB	Active	35	OK
16	Ferritin	Active	287	OK
7, 8	Folate II	Active	72	OK
18	Free PSA	Active	44	OK
5	FSH	Active	68	OK
13	FT3_6	Active	73	Disabled
22	FT4_6M	Active	478	OK

Why this Example ?


Product Information on (September-10-2019)
New assay file, new reagent new calibrator (6 point)

- Recalibrated using the newly arrived calibrator.

Single/Company control Lot # 01141U100



Bio-Rad Control Lot # 40970



- Problem solved, Q.C results within the acceptable range.
- Resumed testing patient samples.

Summary

4- Case # 4 (Free T3)

- New 6 point Calibrator started
- 2 lot # Bio-Rad Q.C outside range
- **Independent Q.C picked up the problem**
- Company Q.C. low, but in range
- C.A.P Proficiency test out
- **Stopped** the Assay.
- Problem solved with new Calibrator

Take Home Message

➤ Medical Technologists

Never recalculate / re-establish Q.C range **unless** there is a very good justification **or** when it is due.

➤ Companies

Educate clients about your products.

➤ Laboratories

Use Independent Control

Establish your Q.C Range

Participate in Monthly Q.C Peer Comparison.

➤ National Guard Hospital

Jeddah, Chemistry staff, Thank You.

Thank You

from

